



Ref. Certif. No.

DE 3 - 4147

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEMESYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC**CB TEST CERTIFICATE
CERTIFICAT D'ESSAI OC**Product
ProduitMedical Electrical Systems
Therapeutic ExerciserName and address of the applicant
Nom et adresse du demandeurmedica-Medizintechnik GmbH
Blumenweg 8
88454 Hochdorf, GERMANYName and address of the manufacturer
Nom et adresse du fabricantmedica-Medizintechnik GmbH, Blumenweg 8, 88454 Hochdorf,
GERMANYName and address of the factory
Nom et adresse de l'usinemedica-Medizintechnik GmbH, Blumenweg 8, 88454 Hochdorf,
GERMANYRating and principal characteristics
Valeurs nominales et caractéristiques principales

Rated voltage:	115 VAC	230 VAC
Rated power:	130 VA	
Rated frequency:	50/60 Hz	
Protection class:	I	
Degree of protection against electric shock:	Type B	

Model/type Ref.
Ref. de type

THERA-live

Additional information (if necessary)
Information complémentaire (si nécessaire)(with the exception of: Clause 52.1 - Programmable Electronic
Systems)A sample of the product was tested and found
to be in conformity with

IEC 60601-1/A2:1995

Un échantillon de ce produit a été essayé et a été
considéré conforme à laas shown in the Test Report Ref. No.
which form part of this certificate
comme indiqué dans le Rapport d'essais numéro
de référence qui constitue une partie de ce
certificatTÜV Product Service
052-080353-000This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**Department: MGE
Date: 2004-11-11
CB 04 11 17939 001**TÜV**
PRODUCT SERVICE

TÜV PRODUCT SERVICE GMBH · Certification Body · Ridlerstrasse 65 · D-80339 München